

LIST  
OF  
THE MAJOR MEDICAL EQUIPMENT  
IN D-BLOCK, UTH

As of 1992

February 22, 1993

UTH Medical Equipment Service Department  
JICA Zambia Infectious Diseases Project

KEY

STATE

- (a) Good Cond.----- This reflects on Equipment in working state.
- (b) Working----- For Equipment is functional state but not fully utilized due to some reason.
- (c) Not working--- Equipment not working due to faults.

COMMENTS

- (a) Never Used----- Equipment which has never been put to use since it came.
- (b) Not Used----- Equipment put out of use due to reasons given.
- (c) Seldom Used----- Equipment not fully utilized as much as it should be.
- (d) Fully Utilized-- Equipment utilized to maximum benefit. This also includes Equipment put in vital areas intended to be used as back-up units e. g. stand-by compressor units
- (e) Lack of manpower to use- There is no adequate manpower to utilize this category of Equipment so far.
- (f) Inadequate sample----- No adequate samples for routine samples due to lack of co-ordination between the wards and the laboratory.
- (g) Inadequate demand----- There is no demand to use Equipment from the wards although there are qualified staff.

US PRICE

The price in Yen is equivalent to dollars during the year of delivery.

NO.	LOCATION	TOTAL MACHINES	WORKING	NOT WORKING	PAGE
1.	EMERGENCY TEST LABORATORY	27	27	0	1~2
2.	SURGICAL PATHOL. LABORATORY	13	12	1	2~3
3.	AUTOPSY ROOM	16	11	5	3~4
4.	PATHOLOGIST DOCTOR'S ROOM	4	4	0	4
5.	THEATRE SURGERY ROOM 1	10	10	0	4~5
6.	THEATRE SURGERY ROOM 2	11	11	0	5
7.	THEATRE ENDOSCOPY ROOM	4	4	0	6
8.	THEATRE CONTROL ROOM	7	7	0	6
9.	THEATRE T. S. U.	10	9	1	7
10.	THEATRE INSTRUMENT STORE ROOM	5	4	1	7
11.	THEATRE RECOVERY ROOM	4	4	0	8
12.	D-11 NURSE OFFICE	9	7	2	8
13.	N. I. C. U. -1	19	19	0	9
14.	N. I. C. U. -2	16	16	0	10
15.	N. I. C. U. -3	12	12	0	11
16.	D-14-1	2	2	0	11
17.	D-14-2	2	2	0	12
18.	D-12-1	21	21	0	13
19.	D-12-2	14	13	1	13~14
20.	D-11 INSTRUMENT STORE ROOM	1	1	0	14
21.	D-11 MILK ROOM	5	5	0	14

NO.	LOCATION	TOTAL MACHINES	WORKING	NOT WORKING	PAGE
22.	D01 NURSE OFFICE	7	6	1	15
23.	D0-1	20	20	0	15~16
24.	D0-1 INSTRUMENT STORE ROOM	2	2	0	16
25.	PHYSICAL EXAMINATION LABORATORY	7	6	1	16
26.	X-RAY ROOM 1	1	1	0	17
27.	X-RAY ROOM 2	1	1	0	17
28.	X-RAY SECRETARY ROOM	3	3	0	17
29.	X-RAY STORE ROOM	23	20	3	18~19
30.	OUT PATIENT DEPARTMENT	8	8	0	19
31.	JAPANESE DOCTOR'S ROOM	11	11	0	19~20
32.	MEDICAL EQUIPMENT SERVICE DEPARTMENT	4	1	3	20
33.	STORE ROOM	25	22	3	21~22
34.	MEDICAL ILLUSTRATION	8	7	1	23
35.	LABOUR WARD	3	3	0	23
36.	THEATRE PHASE 3	2	2	0	24
37.	BLOOD BANK	1	1	0	24
38.	NUCLEAR PHYSICS	1	1	0	24
39.	I. C. U.	2	2	0	24
40.	MAIN LABORATORY	1	1	0	24
TOTAL		342	325 (95%)	17 (5%)	

LOCATION: Emergency Test Laboratory

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Hemoglobin Meter	ERMA Hb-240	671	Good Cond.	Fully Utilized	/	
2.	Hemoglobin Meter	ERMA Hb-240	205444	Good Cond.	Not Used, Lack of manpower to use	/	
3.	Hemoglobin Meter	ERMA Hb-240	669	Good Cond.	Not Used, Lack of manpower to use	/	
4.	Hemoglobin Meter	ERMA Hb-240	670	Good Cond.	Not Used, Lack of manpower to use	/	
5.	Particle Counter	ERMA PC-820	8212	Good Cond.	Fully Utilized	/1983	
	Auto Dilutor	ERMA AD-70	3452	Good Cond.	Fully Utilized	/1983	
6.	Bilometer	MOCHIDA MEB-552	D-10510	Good Cond.	Fully Utilized	/1982	
7.	Refractometer	ATAGO SPR-T2	8205	Good Cond.	Not Used, Inadequate demand	/1981	
8.	Refractometer	ATAGO SPR-T2	8007	Good Cond.	Not Used, Inadequate demand	/1981	
9.	Water Distiller	YAMATO WA-52	171014	Good Cond.	Fully Utilized	/1982	
10.	Water Distiller	YAMATO WA-22	167006	Good Cond.	Fully Utilized	/1982	
11.	Clini Bath	SAKURA KF-2	4204386	Good Cond.	Seldom Used	/1982	
12.	Refrigerated Centrifuge	HITACHI 18PR-52	62252	Good Cond.	Seldom Used	/	
13.	Multi Mixer	IUCHI --	--	Good Cond.	Seldom Used	9/1989	
14.	Discussion Micro Scope	OLYMPUS BH-2	65884	Good Cond.	Fully Utilized	/	
	Exposure Control Unit	OLYMPUS PM-CBAD	204034	Good Cond.	Seldom Used	/	
	Medical Camera	OLYMPUS C-35AD	206089	Good Cond.	Seldom Used	/	
15.	Hematocrit Centrifuge	KOKUSAN H-25F3	80346	Good Cond.	Fully Utilized	/	
16.	Video Chemistry	AMCO SYSTEM 600	6511025	Working	Fully Utilized, Not complicated on prog. 12/1990	/	
17.	Refrigerator	TOSHIBA GR-451EG	000671	Good Cond.	Fully Utilized	/	
18.	Refrigerator	TOSHIBA GR-451EG	000674	Good Cond.	Fully Utilized	/	

To be continued

19. O. H. P. Projector	ELMO	HP-2450	61985	Good Cond.	Fully Utilized	/
20. Discussion Micro Scope	OLYMPUS	BH-2	00345	Good Cond.	Fully Utilized	/
21. Manual WBC Counter	ERMA	12KEY	3379	Good Cond.	Fully Utilized	/1987
22. Manual WBC Counter	ERMA	12KEY	3382	Good Cond.	Fully Utilized	/1987
23. <u>Electro Phoresis Machine</u>	COSMO	MODEL71	--	Working.	Seldom Used	/
24. <u>Bench Centrifuge</u>	JECONS	BS4000	8611336	Good Cond.	Not Used	/
25. <u>Colorimeter</u>	CORNING	253	2674	Good Cond.	Not Used	/
26. Bilmeter	MOCHIDA	MED332	D-10506	Good Cond.	Not Used	/1982
27. Bilmeter	MOCHIDA	MED332	D-10690	Good Cond.	Not Used	/1982

LOCATION: Surgical Pathol. Laboratory

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	USPRICE
1.	-40°C Deep Freezer	SANYO MDF-400	800158	Good Cond.	Fully Utilized	/1983	
2.	-40°C Deep Freezer	SANYO MDF-400	800157	Not working.	Faulty of compressor	/1983	
3.	Deep Freezer	Ocean 410	593776	Good Cond.	Fully Utilized	/	
4.	Incubator	SAKURA IF-2B	4205086	Good Cond.	Not Used, Inadequate demand	/	
5.	Flame Photometer	Lyte Tek 865	96147	Working.	Not Used, accessories missing	/	
6.	Penetration Mixer	SAKURA VP-10B	506073	Good Cond.	Fully Utilized	/	
7.	Video Microscope System						
	Color Video Camera	JVC BY-110CH	06151647	Good Cond.	Never Used, Lack of manpower to use	/	1,400
	Color Video Monitor	JVC AV-MT15	14050563	Good Cond.	Never Used, Lack of manpower to use	/	450
	Camera Adaptor	JVC AA-C11	13950546	Good Cond.	Never Used, Lack of manpower to use	/	400
	Microscope for Above	TIYODA --	63858	Good Cond.	Never Used, Lack of manpower to use	/	800

To be continued

8. Refrigerated Centrifuge	HITACHI	05PR-22	72807	Good Cond.	Not Used, Lack of manpower to use	/
9. Centrifuge	TONY SEIKO	LC06-SP	79185	Good Cond.	Not Used, Lack of manpower to use	/
10. Plate Mixer	KAGAYAKI	MW-2N	0653	Good Cond.	Seldom Used	/
11. Medical Freezer	SANYO	MDF-535D	0020J71	Good Cond.	Fully Utilized	12/1990
12. Draft Chamber	NOYS	--	--	Good Cond.	Fully Utilized	/
13. Fluorescent Microscope	OLYMPUS	BH	661194	Good Cond.	Fully Utilized	/

LOCATION: Autopsy Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DERIVERY DATE	USPRICE
1.	Mortuary Refrigerator	SAKURA MRS-1	4204373	Not Working.	Life-span expired	/1982	
2.	Microtom Knife Sharpner	SAKURA MN-72E	4408116	Good Cond.	Never Used, Lack of manpower to use	/1982	
3.	Blood Gas Analyzer	RADIOMETER ABL-2	56R31N36	Not Working.	Life-span Expired	/1981	
4.	Hemoximeter	RADIOMETER OSM-2	47R25N38	Not Working.		/	
5.	Platelet Centrifuge	ERMA PLC-220	2191	Good Cond.	Not Used, Inadequate demand	/1983	
6.	Platelet Centrifuge	ERMA PLC-220	2210	Good Cond.	Not used, Inadequate demand	/1983	
7.	Platelet Counter	ERMA ACM-403	4351	Good Cond.	Not Used, Inadequate demand	/1983	
8.	Platelet Counter	ERMA ACM-403	4401	Good Cond.	Not Used, Inadequate demand	/1983	
9.	Hematocrit Centrifuge	KUBOTA KH-120A	--	Working.	Not Used, Inadequate demand	/	
10.	Centrifuge	HITACHI SCT-58	87497	Good Cond.	Not Used, Lack of manpower to use	/	
11.	Ultra Hist Dyer	SAKURA RSH-100II	4201213	Good Cond.	Never Used, Lack of manpower to use	/1982	
12.	Ultra Hist Dyer	SAKURA RSH-100II	4201273	Good Cond.	Never Used, Lack of manpower to use	/1982	
13.	Blood Gas Analyzer	RADIOMETER ABL-2	56R38N29	Good Cond.	Not Used, Lack of manpower to use	/1985	
14.	Spectro Photometer	HITACHI 200-20	--	Good Cond.	Not Used, Inadequate demand	/1982	

To be continued

15. Flame Photometer	HITACHI 775	5733-20	Not Working. Waiting. CPU assembly	/1982
Air Compressor for above	HITACHI SC-62	62001	Good Cond.	/1982
16. Chemtek	AMCO SYSTEM600	6511025	Not Working. Life-span Expired	/1982

LOCATION: Pathologist Doctor's Office

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	USPRICE
1.	16mm Projector	ERMO 16-AAR	--	Good Cond.	Seldom Used	/1981	
2.	Slide Projector	ERMO H-300	80968	Good Cond.	Seldom Used	/1981	
3.	Microscope	TIYODA R21trIII	62342	Working.	Not Used, No bulb	/	
4.	Microscope	TIYODA R21trIII	62352	Working.	Not Used, No bulb	/	

LOCATION: Theatre, Surgery Room 1

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	USPRICE
1.	Operating Lamp	YAMADA KY-9856EL	3KY-604	Good Cond.	Fully Utilized, Focus cont. knob broken	/1982	
2.	Operating Lamp	YAMADA KY56AS	50K-541	Good Cond.	Fully Utilized	/1982	
	Power Pack for Above	6001-L	SZ-62A	Good Cond.	Fully Utilized	/	
3.	Operating Table	MIZUHO SPL-330N	80521846	Good Cond.	Fully Utilized	/1982	
4.	Anaesthesia Apparatus	ACOMA PH-3S	11759	Working.	Fully Utilized, Waiting, bulve	/1982	
5.	Ventilator	ACOMA ARP-850	850	Good Cond.	Fully Utilized	/1982	
6.	Pharmaceutical Refrigerator	SANYO MPR-110	820322	Good Cond.	Fully Utilized	/1982	
7.	Infusion Pump	IVAC 531	73622	Good Cond.	Fully Utilized	/1982	
8.	Film Illuminator	MIWA MQ-HU31A	693	Good Cond.	Seldom Used	/1982	

To be continued



9. Blood Warmer TORAY W-2000 818044 Good Cond. Fully Utilized /  
 10. Electro Surgery ERBOTOM 400TC 005-E-2361 Good Cond. Fully Utilized /

LOCATION: Theatre, Surgery Room 2

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DERIVERY DATE	US PRICE
1.	Operating Lamp	YAMADA KY-56AS	50K-540	Good Cond.	Fully Utilized	/1982	
2.	Operating Lamp	YAMADA KY-56AS	50K-539	Good Cond.	Fully Utilized	/	
3.	Operating Lamp	YAMADA KY-9856EL	2KY-605	Good Cond.	Fully Utilized	/1982	
	Power Pack for Above	6001-L	SZ-61A	Good Cond.	Fully Utilized	/	
4.	Operating Table	MIZUHO SPL-330N	80521853	Good Cond.	Fully Utilized	/1982	
5.	Anaesthesia Apparatus	ACOMA PH-3S	11760	Working.	Fully Utilized, Waiting bulve	/1982	
6.	Suction Machine	ATOM D-80	895	Good Cond.	Fully Utilized	/1981	700
7.	Blood Warmer	TORAY W-2000	818051	Good Cond.	Fully Utilized	/1982	
8.	Film Illuminator	MIWA MS-HUBA	692	Good Cond.	Seldom Used	/1982	
9.	Life Monitor	NIHONKOHDEN OEC-5200	20406	Good Cond.	Fully Utilized	/	
	Recorder	NIHONKOHDEN WT-311R	20969K	Good Cond.	Not Used, inadequate demand	/	
10.	Life monitor	NIHONKOHDEN OMP-3200	20909	Good Cond.	Fully Utilized	/	
	Recorder	NIHONKOHDEN WT-311R	20988K	Good Cond.	Not Used, inadequate demand	/	
11.	Electrical Surgery	VallayLab SSE-3B H5H4808B-1		Good Cond.	Fully Utilized	/	

LOCATION: Theatre, Endoscopy Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DERIVERY DATE	US PRICE
1.	Operating Table	MIZUHO SPL-330N	80521866	Good Cond.	Not Used, Lack of manpower to use	/1982	
2.	Photographing Table	SAKURA UP-R	R4601716	Good Cond.	Never Used, Inadequate demand	/	
3.	Operating Lamp	YAMADA KY-56AS	5DK-539	Good Cond.	Not Used, Lack of manpower to use	/1982	
4.	Operating Lamp	YAMADA KY-96	4K-249	Working.	Not Used, Faulty of battery	/1982	

LOCATION: Theatre, Control Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DERIVERY DATE	US PRICE
1.	Infant Humidifier	ATOM 105	3786	Working.	Not Used, Plug removed	/1982	950
2.	Infant Humidifier	ATOM 105	4987	Good Cond.	Not Used, Inadequate demand	/1982	950
3.	Life Monitor Recorder	NIHONKOHDEN OEC-3200	20910	Good Cond.	Not Used, Inadequate demand	/1982	5,100
		NIHONKOHDEN WT-311R	20989K	Good Cond.	Not Used	/1982	
4.	Life Monitor	NIHONKOHDEN OEC-3200	20911	Good Cond.	Not Used, Inadequate demand	/1982	4,600
5.	Infusion Pump	IVAC 530	53421	Working.	Not Used, Plug removed	/1982	
6.	Infusion Pump	IVAC 530	73223	Good Cond.	Not Used, Inadequate demand	/1982	
7.	Infusion Pump	IVAC 530	73317	Working.	Not Used, Plug removed	/1982	

LOCATION: Theatre, T.S.S.U.

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DERIVERY DATE	US PRICE
1.	Hot Air Sterilizer	SAKURA HE-11	4512909	Good Cond.	Seldom Used	/	
2.	Hot Air Sterilizer	SAKURA HE-11	4512910	Working.	Not Used, 13A fuse removed.	/	
3.	Hot Air Sterilizer	SAKURA HE-11	--	Good Cond.	Not Used	/	
4.	Boiling Sterilizer	SAKURA FA-360BE	--	Not working.	Waiting heater element	/1982	
5.	Steam Sterilizer	SAKURA TF-31	--	Good Cond.	Not Used	/1982	
6.	Steam Sterilizer	SAKURA FBA-6S2PC	--	Good Cond.	Fully Utilized	/1982	
7.	Electric Sterilizer	SAKURA BA-2E	042072707	Good Cond.	Not Used	/	
8.	Steam Sterilizer	SAKURA SP-203	4206419	Good Cond.	Not Used, Inadequate demand	/	
9.	Difbrillator	NIHONKOHDEN TEC-3500	20484	Good Cond.	Seldom Used	/	
10.	Life Monitor	NIHONKOHDEN OEC-3500	20391	Good Cond.	Seldom Used	/	

LOCATION: Theatre, Instrument Store Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DERIVERY DATE	US PRICE
1.	Pharmaceutical Refrigerator	SANYO MPR-110	820325	Good Cond.	Fully Utilized	/1982	
2.	Anaesthesia Apparatus	ACOMA PH-3	10736	Good Cond.	Not Used	/1982	
3.	Cold Light Supply	OLYMPUS CLE-4U	2406955	Good Cond.	Never Used, Lack of manpower to use	/1982	
	Gastroscope	OLYMPUS BF, 4B2	--	Good Cond.	Not Used	/1982	
4.	Cold Light Supply	OLYMPUS CLE-F	220470	Good Cond.	Never Used, Lack of manpower to use	/1982	
	Gastroscope	OLYMPUS PGF, S2	--	Good Cond.	Not Used	/1982	
5.	Cold Light Supply	AESUCLAP OP-900	886234	Not Working.	Lamp and lens broken	/	
	To be continued						

LOCATION: Theatre, Recovery Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Incubator	ATOM V-80MC	8034014, 8038223	Good Cond.	Furry Utilized	/1988	4,600
2.	Infant Humidifier	ATOM 105	3785	Good Cond.	Furry Utilized	/1982	950
3.	Body Warmer	GAYMAR TP-220	0006	Good cond.	Fully Utilized	/1982	
4.	Suction Machine	ATOM D-58	9031895	Good cond.	Fully Utilized	/	

LOCATION: D-11, Nurse Office

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Film Illuminator	MIWA MS-T2F	687	Good Cond.	Seldom Used	/1982	
2.	Ice Maker	SANYO SIM-EJ00W	100168	Not Working.	Life-span expired	/1982	
3.	Hematocrit Centrifuge	HITACHI MC-202	78423	Good Cond.	Fully Utilized	/	
4.	Bilometer	MOCHIDA MER-552	D10509	Good Cond.	Fully Utilized	/1982	
5.	Pharmaceutical Refrigerator	SANYO MPR-110	820324	Good Cond.	Fully Utilized	/1982	
6.	Refrigerator	TOSHIBA GR-262E(G)	001024	Good Cond.	Fully Utilized	/	
7.	Bilirubin Meter	MINORUTA 54B-743	112026	Good Cond.	Seldom Used	/1991	
8.	Blood Gas Analyzer	CHIBACORNING 278	2410	Not Working.	Neccessary to change all of Electrode	/1990	66,000
9.	Refractometer	ATAGO SPR-T2		Good Cond.	Not Used	/1981	

LOCATION: N. I. C. U. -1

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Respirator	BOURNS BP-202	10518	Good Cond.	Fully Utilized	/1983	
2.	Infant Humidifier	BOURNS LS-460	--	Good Cond.	Fully Utilized	/1983	
3.	Infant Respirator	BOURNS BP-202	10627	Good Cond.	Fully Utilized	/1983	
4.	Humidifier	BOURNS --	--	Good Cond.	Fully Utilized	/1983	
5.	Infant Warmer	ATOM V-3200	1026	Good Cond.	Fully Utilized	/1982	5,100
6.	Infant Incubator	ATOM V-80MC	1225, 02573	Good Cond.	Fully Utilized	/1982	3,450
7.	Infant Incubator	ATOM V-80MC	1240, 02569	Good Cond.	Fully Utilized	/1982	3,450
8.	Infant Incubator	ATOM V-80MC	1233, 02570	Good Cond.	Fully Utilized	/1982	3,450
9.	Infant Incubator	ATOM V-80MC	1246, 02539	Good Cond.	Fully Utilized	/1982	3,450
10.	Infant Incubator	ATOM V-80MC	1224, 02538	Good Cond.	Fully Utilized	/1982	3,450
11.	Infant Incubator	ATOM V-80MC	8034019, 8038228	Good Cond.	Fully Utilized	/1988	4,600
12.	Oxygen Analyzer	ATOM OX-55	1045	Good Cond.	Fully Utilized	/1982	700
13.	Life Monitor	NIHONKOHDEN OMP-3221	20693	Good Cond.	Fully Utilized	/1982	4,600
14.	Infusion Pump	IVAC 531	73655	Good Cond.	Seldom Used	/1982	
15.	Pulse Oximeter	NIHONKOHDEN OLV-1200K	20018	Good Cond.	Fully Utilized	/1991	
16.	Infant Incubator	ATOM V-80MC	02533	Good Cond.	Fully Utilized	/	
17.	Infant Warmer	ATOM V-3200	1017	Good Cond.	Fully Utilized	/	
18.	Infant Warmer	ATOM V-3200	822	Good Cond.	Fully Utilized	/	
19.	Phototherapy Unit	ATOM PIT-150st	319	Good Cond.	Fully Utilized	/	

LOCATION: N. J. C. U. -2

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Respirator	BOURNS BP-202	10501	Good Cond.	Fully Utilized	/1983	
2.	Infant Humidifier	BOURNS LS-460	--	Good Cond.	Fully Utilized	/1983	
3.	Infant Respirator	BOURNS BP-202	10509	Good Cond.	Fully Utilized	/1983	
4.	Infant Humidifier	BOURNS LS-460	--	Good Cond.	Fully Utilized	/1983	
5.	Life Monitor	NIHONKOHDEN OMP-3221	20700	Good Cond.	Fully Utilized	/1982	4,600
6.	Infant Warmer	ATOM V-3200	0480	Good Cond.	Fully Utilized	/1982	1,100
7.	Infant Incubator	ATOM V-80MC	1247, 02529	Good Cond.	Fully Utilized	/1982	3,450
8.	Infant Incubator	ATOM V-80MC	00150, 02572	Good Cond.	Fully Utilized	/1982	3,450
9.	Infant Incubator	ATOM V-80MC	00146, 02536	Good Cond.	Fully Utilized	/1982	3,450
10.	Infant Incubator	ATOM V-80MC	0341, 1258	Good Cond.	Fully Utilized	/1982	4,600
11.	Infant Incubator	ATOM V-80MC	S034018, 8038227	Good Cond.	Fully Utilized	/1988	4,600
12.	Infant Incubator	ATOM V-80MC	00149, 02531	Good Cond.	Fully Utilized	/1982	3,450
13.	Phototherapy Unit	ATOM PIT-120ST	--	Good Cond.	Fully Utilized	/1982	1,100
14.	Phototherapy Unit	ATOM PIT-120ST	1385	Good Cond.	Fully Utilized	/1982	1,100
15.	Infant Respirator	BOURNS BP-202	10504	Good Cond.	Fully Utilized	/	
16.	Infant Incubator	ATOM V-80MC	02540	Good Cond.	Fully Utilized	/	

LOCATION: N. I. C. U. -3

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Ventilator	BOURNS BP-202	10611	Good Cond.	Fully Utilized	/	
2.	Ventilation Monitor	BOURNS LS-162	3043	Good Cond.	Fully Utilized	/	
3.	Infant Humidifier	BOURNS --	--	Good Cond.	Fully Utilized	/	
4.	Infant Warmer	ATOM V-3200	1137	Good Cond.	Fully Utilized	/1982	4,150
5.	Infant Incubator	ATOM V-80MC	1219, 02527	Good Cond.	Fully Utilized	/1982	3,450
6.	Infant Incubator	ATOM V-80MC	1255, 02550	Good Cond.	Fully Utilized	/1982	3,450
7.	Infant Incubator	ATOM V-80MC	1364, 02552	Good Cond.	Fully Utilized	/1982	3,450
8.	Infant Incubator	ATOM V-80MC	00338, 1245	Good Cond.	Fully Utilized	/1982	3,450
9.	Infant Incubator	ATOM V-80MC	8034017, 8038224	Good Cond.	Fully Utilized	/1988	4,600
10.	Infant Incubator	ATOM V-80MC	8038226, 8038226	Good Cond.	Fully Utilized	/1988	4,600
11.	Life monitor.	NIHONKOHDEN OMP-3221	20699	Good Cond.	Fully Utilized	/1982	4,600
12.	Life Monitor	NIHONKOHDEN OMP-3221	20701	Good Cond.	Not Used, Inadequate demand	/1982	5,100
	Recorder	NIHONKOHDEN WT-311R	20973K	Good Cond.	Not Used, Inadequate demand	/1982	

LOCATION: D-14-1

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Incubator	ATOM V-80MC	1229, 02574	Good Cond.	Fully Utilized	/1982	3,450
2.	Phototherapy Unit	ATOM PIT-220	9030087	Good Cond.	Fully Utilized	6/1990	1,600

LOCATION: D-14-2

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Incubator	ATOM V-80MC	1220, 02528	Good Cond.	Fully Utilized	/1982	3,450
2.	Phototherapy Unit	ATOM PIT-220	9050088	Good Cond.	Fully Utilized	6/1990	1,600

LOCATION: D-12-1

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1. Portable X-ray Equipment							
	Main Body	SHIMADZU M0125L-30	6231861	Good Cond.	Fully Utilized	/	
	Corimeter	SHIMADZU R-20MC	62823041	Good Cond.	Fully Utilized	/	
	X-ray Tube	SHIMADZU --	280833	Good Cond.	Fully Utilized	/	
2.	Air Compressor for Buck-up	--	MACS-310	550353	Good Cond.	Never Used, Inadequate demand	/
3.	Air Compressor for Buck-up	AMCO MC-30	2110286	Good Cond.	Good Cond.	Not Used, Air connector missing	/1983
4.	Air Compressor for Buck-up	AMCO MC-30	2110287	Good Cond.	Good Cond.	Not Used, Air connector missing	/1983
5.	Air Compressor for Buck-up	AMCO MC-30	2110288	Good Cond.	Good Cond.	Not Used, Air connector missing	/1983
6.	Infant Incubator	ATOM V-80MC	1256, 02567	Good Cond.	Fully Utilized	/1982	3,450
7.	Infant Incubator	ATOM V-80MC	1221, 02576	Good Cond.	Fully Utilized	/1982	3,450
8.	Infant Incubator	ATOM V-80MC	00151, 02526	Good Cond.	Fully Utilized	/1982	3,450
9.	Infant Incubator	ATOM V-80MC	1222, 02534	Good Cond.	Fully Utilized	/1982	3,450
10.	Infant Incubator	ATOM V-80MC	00148, 00340	Good Cond.	Fully Utilized	/1982	3,450
11.	Infant Incubator	ATOM V-80MC	1228, 02535	Good Cond.	Fully Utilized	/1982	3,450
12.	Infant Incubator	ATOM V-80MC	1232, 00337	Good Cond.	Fully Utilized	/1988	3,450

To be continued



13. Infant Incubator	ATOM V-80MC	1242, 2571	Good Cond.	Fully Utilized	/1988	3,450
14. Infant Incubator	ATOM V-80MC	1230, 02541	Good Cond.	Fully Utilized	/1988	3,450
15. Infant Incubator	ATOM V-80MC	8034023, 8038232	Good Cond.	Fully Utilized	/1988	4,500
16. Infant Incubator	ATOM V-80MC	1231, 02552	Good Cond.	Fully Utilized	/1982	3,450
17. Infant Incubator	ATOM V-80MC	1244, 02575	Good Cond.	Fully Utilized	/1982	3,450
18. Infant Incubator	ATOM V-80MC	319-8038230	Good Cond.	Fully Utilized	/1982	3,450
19. Transport Incubator	OHIO MEDI. --	AKJM00081	Working	Fully Utilized	/1981	
Control Unit	OHIO MEDI. --	AKHM00026	Good Cond.	Fully Utilized		
20. Phototherapy Unit	ATOM PIT-120st.	1389	Good Cond.	Fully Utilized	/1981	1,100
21. Infant Warmer	ATOM V-3200	1190	Good Cond.	Fully Utilized	/	

LOCATION: D-12-2

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infusion Pump	IVAC 530	53408	Good Cond.	Not Used, Inadequate demand	/1982	
2.	Infusion Pump	IVAC 531	73534	Good Cond.	Not Used, Inadequate demand	/1982	
3.	Infusion Pump	IVAC 531	73541	Good Cond.	Not Used, Inadequate demand	/1982	
4.	Infusion Pump	IVAC 531	73682	Good Cond.	Not Used, Inadequate demand	/1982	
5.	Infusion Pump	IVAC 531	73656	Good Cond.	Not Used, Inadequate demand	/1982	
6.	Infant Humidifier	ATOM 105	4944	Good Cond.	Fully Utilized	/1982	950
7.	Infant Humidifier	ATOM 105	5033	Good Cond.	Fully Utilized	/1982	950
8.	Infant Incubator	ATOM V-80MC	1227, 02537	Good Cond.	Fully Utilized	/1982	3,450
9.	Infant Incubator	ATOM V-80MC	1241, 2558	Good Cond.	Fully Utilized	/1982	3,450
10.	Infant Incubator	ATOM V-80MC	1257, 00339	Good Cond.	Fully Utilized	/1982	3,450

To be continued

11. Infant Incubator	ATOM V-80MC	8034020, 8038229	Good Cond.	Fully Utilized	/1982	3,450
12. Infant Incubator	ATOM V-80MC	8034016, 8038225	Good Cond.	Fully Utilized	/1988	4,600
13. Infant Incubator	ATOM V-80MC	8034022, 8038231	Good Cond.	Fully Utilized	/1988	4,600
14. Stericabinet	TOYODA TFA-10	--	Not Working.	Can not be repaired, miss used	/1988	2,000

LOCATION: D-11, Instrument Store Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
-----	-------------------	---------------	------------	-------	----------	---------------	----------

1.	Dextrometer	SANKYO 5564	208782	Good Cond.	Not Used, Lack of manpower to use	/	
----	-------------	-------------	--------	------------	-----------------------------------	---	--

LOCATION: D-11, Milk Room

1.	Steam Sterilizer	SAKURA FYA-50S	--	Good Cond.	Fully Utilized	/	
2.	Pharmaceutical Refrigerator	SANYO MPR-500	800159	Good Cond.	Fully Utilized	/1981	
3.	Steam Bottle Sterilizer	SANKO --	--	Good Cond.	Fully Utilized	/1981	
4.	Nursing Bottle Warmer	ATOM NS-110	1004	Good Cond.	Fully Utilized	/1981	800
5.	Nursing Bottle Warmer	ATOM NS-110	826	Good Cond.	Fully Utilized	/1981	800

LOCATION: DO-1, Nurse Office

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	X-ray Film Illuminator	MIWA MS-T21	686	Good Cond.	Seldom Used	/	
2.	Blood Warmer	TORAY W-2000	818123	Good Cond.	Seldom Used	/	
3.	Life Monitor	NIHONKOHDEN OMP-S221	20725	Good Cond.	Seldom Used	/1982	4,600
4.	Ice Maker	SANYO SIM-100W	100167	Not Working	Life-span expired	/1982	
5.	Nursing Bottle Warmer	ATOM NS-110	1005	Good Cond.	Fully Utilized	/1982	800
6.	Pharmaceutical Refrigerator	SANYO MPR-110	80323	Good Cond.	Fully Utilized	/	
7.	1ch ECG Machine	NIHONKOHDEN ECG-5101	21094	Good Cond.	Fully Utilized	/	

LOCATION: DO-1

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Incubator	ATOM V-55L	2236	Good Cond.	Fully Utilized	/	
2.	Infant Incubator	ATOM V-55L	2237	Good Cond.	Fully Utilized	/	
3.	Infant Incubator	ATOM V-55L	2238	Good Cond.	Fully Utilized	/	
4.	Infant Incubator	ATOM V-55L	2239	Good Cond.	Fully Utilized	/	
5.	Phototherapy Unit	ATOM PIT-220	9030086	Good Cond.	Not Used, Inadequate demand	6/1990	1,600
6.	Paediatric Ventilator	BENNETT PR-2	--	Good Cond.	Not Used, Lack of manpower to use	/	
7.	Suction Machine	ATOM D-34	--	Good Cond.	Fully Utilized	/1982	700
8.	Pharmaceutical Refrigerator	SANYO MPR-110	820323	Good Cond.	Fully Utilized	/1982	
9.	Portable X-ray Equipment	TOSHIBA		Good Cond.	Seldom Used	/1988	
	X-ray Tube	TOSHIBA D-1201	12645	Good Cond.	Seldom Used	/	
	Main Body	TOSHIBA TOREX	3517702	Good Cond.	Seldom Used	/	
	Control Unit	TOSHIBA TR-80A	8538459	Good Cond.	Seldom Used	/	
10.	Back-up Compressor	MACS-310	--	Good Cond.	Never Used, No connected to 3 $\phi$ 380V	/	
11.	Transport Incubator	OHIO AIR VAC	AKJH00294	Good Cond.	Seldom Used	/1982	
	Controller		AKHH00208	Good Cond.	Seldom Used	/	
	To be continued						

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
12.	Infant Humidifier	ATOM 105	5032	Good Cond.	Not Used, Lack of manpower to use	/1982	950
13.	Infant Humidifier	ATOM 105	5034	Good Cond.	Not Used, Lack of manpower to use	/1982	950
14.	Oxygen Tent	ATOM OX-61sc	614982	Good Cond.	Fully Utilized	/	3,400
15.	Oxygen Tent	ATOM OX-61sc	615462	Good Cond.	Fully Utilized	/	3,400
16.	Oxygen Tent	ATOM OX-61sc	615479	Good Cond.	Fully Utilized	/	3,400
17.	Continuous Suction Unit	MIZUHO MSP-210	4708	Good Cond.	Not Used, Inadequate demand	/1982	
18.	Continuous Suction Unit	MIZUHO MSP-210	4709	Good Cond.	Not Used, Inadequate demand	/1982	
19.	Continuous Suction Unit	MIZUHO MSP-210	--	Good Cond.	Not Used, Inadequate demand	/1982	
20.	Infant Ventilator	BABY BIRD 5900	422490	Good Cond.	Not Used.	/	

LOCATION: DO-1, Instrument Store Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Continuous Suction Machine	MIZUHO MSP-210	3187	Good Cond.	Not Used, Inadequate demand	/	
2.	Suction Machine	ATOM D-58	9031894	Good Cond.	Never Used, Spare equipment.	/1990	1,100

LOCATION: Physical Examination Laboratory

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Ultrasound Diagnostic Eq.	SHIMADZU SDL-100A	68400302	Working	Faulty of CPU Assy.	/1982	
2.	Ultrasound Diagnostic Eq.	ALOKA SSD-720	41M1566	Good Cond.	Fully Utilized	/1984	30,000
3.	9ch EEG Machine	NIHONKOHDEN EEG-7209	20532	Not Working	Faulty of preamplifier	/	17,000
4.	3ch ECG Machine	NIHONKOHDEN ECG-5303	20562	Good Cond.	Not Used, Lack of manpower to use	/1982	
5.	Ultrasonnd Diagnostic Eq.	TOSHIBA SSA-90A	F7562987	Working	Fully Utilized, Foot Sw. damaged	/1987	80,000
	Sono printer	TOSHIBA TP-8300	E7520738	Good Cond.	Fully Utilized	/1987	
	Power Stabilizer	MATSUNAGA SUZ-2214	--	Good Cond.	Fully Utilized	/1987	
6.	Ultrasound	SIEMENS EE72115		Good Cond.	Fully Utilized	/	
	Sono printer	MIITSUBISHI IEP61E		Good Cond.	Fully Utilized	/	
7.	3ch ECG Machine	NIHONKOHDEN ECG-5303	20561	Good Cond.	Not Used, Lack of manpower to use	/	

LOCATION: X-ray Room 1

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	X-ray Equipment	SHIMADZU FH-21	62712905	Good Cond.	Fully Utilized	13/02/1984	
	Tube	SHIMADZU R-20	62816208	Good Cond.	Fully Utilized	13/02/1984	
	Control Unit	SHIMADZU ED-150L	6218984	Good Cond.	Fully Utilized	13/02/1984	
		SHIMADZU BF-11	04512322	Good Cond.	Fully Utilized	13/02/1984	
		SHIMADZU BR-10M	6267832	Good Cond.	Fully Utilized	13/02/1984	
		SHIMADZU ED150L	64234606	Good Cond.	Fully Utilized	13/02/1984	
		TRANS PF16	6261084	Good Cond.	Fully Utilized	13/02/1984	

(DATE OF INSTALLATION)

LOCATION: X-ray Room 2

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	X-ray Equipment	SHIMADZU FH-21	62712909	Good Cond.	Fully Utilized	/1984	
	Tube	SHIMADZU R-20	62816244	Good Cond.	Fully Utilized	/1984	
	Tube	SHIMADZU 1A-9/5X	110525	Good Cond.	Fully Utilized	/1984	
	Control Unit	SHIMADZU ED-150L	6218983	Good Cond.	Fully Utilized	/1984	
	X-ray Monitor	SHIMADZU YSM-1591	6251521	Good Cond.	Fully Utilized	/1984	
	X-ray Monitor	MITSUBISHI M-1635C	7156020	Good Cond.	Fully Utilized	/1984	
		MITSUBISHI M-1665C	7301002	Good Cond.	Fully Utilized	/1984	

LOCATION: X-ray Secretary Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Automatic Film Developer	YOKOYAMA SR-90	--	Working.	Fully Utilized, Waiting Pump	/	
2.	X-ray Film Illuminator	MIWA MS-T612S	691	Good Cond.	Fully Utilized	/	
3.	X-ray Film Illuminator	MIWA MS-T612S	--	Good Cond.	Fully Utilized	/	

LOCATION: X-ray Store Room

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Humidifier	BOURNS LS-460	5396	Good Cond.	Not Used, Inadequate demand	/	
2.	Neonatal Monitor	ATOM V-800M	1152	Good Cond.	Not Used, Inadequate demand	/1981	4,600
3.	Neonatal Monitor	ATOM 512	512088398	Good Cond.	Not Used, Inadequate demand	/1981	
4.	Infant Humidifier	ATOM 105	3789	Good Cond.	Not Used, Inadequate demand	/1981	950
5.	Infant Humidifier	ATOM 105	5031	Good Cond.	Not Used, Inadequate demand	/1981	950
6.	Infusion Pump	ATOM R-100	1726	Good Cond.	Not Used, Inadequate demand	/	1,600
7.	Infusion Pump	ATOM R-100	1727	Good Cond.	Not Used, Inadequate demand	/	1,600
8.	Infusion Pump	ATOM R-100	1729	Good Cond.	Not Used, Inadequate demand	/	1,600
9.	Infusion Pump	ATOM R-100	1730	Good Cond.	Not Used, Inadequate demand	/	1,600
10.	Suction Machine	ATOM D-34	73261	Not Working,	Life-span expired	/1981	700
11.	Suction Machine	ATOM D-34	73658	Not Working,	Life-span expired	/1981	700
12.	Hemoshaker	TERUMO ACS-169	8402003	Good Cond.	Never Used	/	
13.	Hemoshaker	TERUMO ACS-169	8402004	Good Cond.	Never Used	/	
14.	Bench Centrifuge	TOMY HC-12A	81108	Not Working.		/	
15.	Vacuum Pump	HITACHI 16VP-D	82012	Good Cond.	Not Used	/	
16.	Universal Washer	MITA RIKA TypeZ	1905	Good Cond.	Never Used	/1981	
17.	Universal Washer	MITA RIKA TypeZ	1904	Good Cond.	Never Used	/1981	
18.	Life Monitor	NIHONKOHDEN OMP-3221	20695	Working.	Not Used, Faulty of transmitter	/1981	4,600
19.	Breast Pump	ATOM BP-80	0769	Good Cond.	Not Used, Inadequate demand	/1981	
20.	ECG Machine	NIHONKOHDEN ECG-5101	21095	Good Cond.	Never Used	/	
21.	TC Oxygen Monitor	RADIOMETER --	280919	Good Cond.	Not Used, Lack of manpower to use	/1981	
	TC Cariblation Unit	RADIOMETER TC-101	55R21N015	Good Cond.	Not Used, Lack of manpower to use	/1981	
	TC Recorder	RADIOMETER TC-R2	0772	Good Cond.	Not Used, Lack of manpower to use	/1981	
22.	TC Oxygen Monitor	RADIOMETER --	280897	Good Cond.	Not Used, Lack of manpower to use	/1981	
	TC Cariblation Unit	RADIOMETER TC-101	283882	Good Cond.	Not Used, Lack of manpower to use	/1981	
	TC Recorder	RADIOMETER TC-R2	0793	Good Cond.	Not Used, Lack of manpower to use	/1981	

To be continued



NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
8.	Copy Machine	XEROX 5012	--	Good Cond.	Fully Utilized	/1991	
9.	Bilirubin Meter	MINORUTA 54B-743	112008	Good Cond.	Fully Utilized	/1990	
10.	Transfusion Pump	ATOM 201	494	Good Cond.	Not Used	/	
11.	Refrigerator	ocean --	--	Good Cond.	Fully Utilized	/1989	
12.	Microscope	OLYMPUS BH-2	206951	Good Cond.	Not Used	/	

LOCATION: Medical Equipment Service Department

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Digital BP Machine	OMRON HEM-50	147449	Good Cond.	Not Used, Lack of manpower to use	/	
2.	Oxygen Analyzer	TELEDYNE TED-100	49433	Not Working	Faulty of sensor	/	
3.	Oxygen Analyzer	ATOM OX-35	834	Not Working	Faulty of Sensor	/	
4.	Oxygen Analyzer	ATOM OX-30	2282	Not Working	Faulty of Sensor	/	



LOCATION: Store Room (Lecture room)

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Neonatal Monitor	ATOM V-800M	1155	Not Working	Mismanagement on repair	/1981	4,600
2.	Electric Surgery	Vallaylab SSE-5B H3E904B-1AF		Not Working	Mismanagement on repair	/	
3.	Infusion Pump	ATOM R-100	1728	Good Cond.	Not Used, Inadequate demand	/	
4.	Infusion Pump	IVAC 531	75740	Good Cond.	Not Used, Inadequate demand	/	
5.	Infant Humidifier	BOURNS LS-460	5380	Good Cond.	Never used, Inadequate demand	/1982	
6.	Suction Machine	ATOM D-80	895	Good Cond.	Not Used, Inadequate demand	/1981	700
7.	Stereo Power amplifier	SONY TA-AX5	514942	Not Working	Big damaged by miss use	/1981	
	Microphone Mixer	SONY MX-670	204289	Good Cond.			
	Dynamic Microphone	SONY F-115A	20141, 15326	Good cond.			
	Speaker System	SONY SS-S51	400902, 400897	Good Cond.			
8.	Suction Machine	ATOM D-58	9031897	Good Cond.	Never Used, Awaiting distribution	/1990	1,200
9.	Suction Machine	ATOM D-58	1012900	Good Cond.	Never Used, Awaiting distribution	6/1991	1,200
10.	Suction Machine	ATOM D-58	1012902	Good Cond.	Never Used, Awaiting distribution	6/1991	1,200
11.	Suction Machine	ATOM D-58	1012901	Good Cond.	Never Used, Awaiting distribution	6/1991	1,200
12.	Bloodless Sphygmomanometer	NIHONKOHDEN MPV-7201	01708	Good Cond.	Never Used, Awaiting distribution	6/1991	4,400
	Printer	NIHONKOHDEN YP-714R	00344	Good Cond.	Never Used, Awaiting distribution	6/1991	1,750
13.	Pulse Oximeter	NIHONKOHDEN OLV-1200K	20019	Good Cond.	Never Used, Awaiting distribution	/1990	7,000
14.	Pulse Oximeter	NIHONKOHDEN OLV-1200K	20017	Good Cond.	Never Used, Awaiting distribution	/1990	7,000
15.	Oxygen Analyzer	AICA OMP-1000	8911040	Good Cond.	Never Used, Awaiting distribution	/1990	1,530
16.	Oxygen Analyzer	AICA OMP-1000	8911044	Good Cond.	Never Used, Awaiting distribution	/1990	1,530
17.	Oxygen Analyzer	AICA OMP-1000	8911043	Good Cond.	Never Used, Awaiting distribution	/1990	1,530

To be continued

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
18.	Multi Mixer	IUCHI	--	Good Cond.	Never Used, Awaiting distribution	/1990	
19.	Controller for ATOM V-80MC	ATOM	1243	Good Cond.	Fully Utilized	/1982	
20.	Controller for ATOM V-80MC	ATOM	1259	Good Cond.	Fully Utilized	/1982	
21.	Controller for ATOM V-80MC	ATOM	00147	Good Cond.	Fully Utilized	/1982	
22.	Controller for ATOM V-80MC	ATOM	1248	Good Cond.	Fully Utilized	/1982	
23.	Controller for ATOM V-80MC	ATOM	1234	Good Cond.	Fully Utilized	/1982	
24.	Controller for OHIO AIR VAC Transport Incubator	AXHR00090		Good Cond.	Fully Utilized	/1990	5,000
25.	Copy Machine	Canon NP-120	30161987	Good Cond.	Fully Utilized	/1985	

THE OTHER WARDS

LOCATION: Medical Illustration

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Microscope	OLYMPUS BH-2	--	Good cond.	Fully Utilized	/1983	
2.	Exposure Control Unit	OLYMPUS PM-CBAD	204044	Good Cond.	Fully Utilized	/1983	
3.	Medical Camera	OLYMPUS C-35AD	205922	Good Cond.	Fully Utilized	/1983	
4.	Portable Video recorder	SONY SL-F1E	200192	Good cond.	Fully Utilized	/1983	
	Battery Charger	SONY BC-20CE	13152	Good cond.	Fully Utilized	/1983	
5.	Video Casset Recorder	SONY SL-C7E	53973	Not Working, Faulty of capstan motor		/1983	
6.	Video Monitor	SONY CVM-200PSE	201319	Good Cond.	Fully utilized	/1983	
7.	Color TV	SONY KV-2032EV	157521	Good cond.	Fully Utilized	/1983	
8.	TV Camera	SONY DXC-1800P	50831	Good Cond.	Fully Utilized	/1983	
	Camera Adaptor	SONY CMA-7CE	14158	Good Cond.	Fully Utilized	/1983	

LOCATION: Labour Ward

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	Infant Incubator	ATOM V-80MC	--	Good Cond.	Fully Utilized	/1981	3,400
2.	Transport Incubator	OHIO AIR VAC	--	Good Cond.	Fully Utilized		
3.	Infant Warmer	ATOM V-3200D	1021171	Good Cond.	Fully Utilized	/1991	5,500

LOCATION: Theatre, Phase III

NO.	NAME OF EQUIPMENT	MAKER & MODEL	SERIAL NO.	STATE	COMMENTS	DELIVERY DATE	US PRICE
1.	X-ray Equipment	SHIMADZU WHA-10	6130250	Good Cond.	Fully Utilized	/	
	X-ray Monitor	MITSUBISHI 903A-11	7534013	Good cond.	Fully Utilized	/	
2.	Life Monitor	NIHONKOHDEN OMP-3221	--	Good Cond.	Fully Utilized	/1982	4,600

LOCATION: Blood Bank

1.	Forma Bio-Freezer	FORMA SCIENTIFIC 51738827-10		Good Cond.	Fully Utilized	/1983	
----	-------------------	------------------------------	--	------------	----------------	-------	--

LOCATION: Nuclear Physics

1.	9CH EEG Machine	NIHONKOHDEN EEG-7209 20531		Good Cond.	Fully Utilized	/	19,000
----	-----------------	----------------------------	--	------------	----------------	---	--------

LOCATION: I.C.U.

1.	Neonatal Monitor	ATOM V-800M	1153	Good Cond.	Fully Utilized	/1981	4,600
2.	Ventilation Unit	BENNETT MA-1B	--	Good Cond.	Fully Utilized	/	

LOCATION: Main Laboratory

1.	Medical Freezer	SANYO MDF-535D	--	Good Cond.	Fully Utilized	/1990	2,900
----	-----------------	----------------	----	------------	----------------	-------	-------







JICA